

London, July 2020

bCUBE 2.0 Quick Start Guide

Hyris Platform guide

Version 1.2.7

Hyris Ltd

Hyris Headquarters

Lower Ground Floor, One George Yard,
EC3V 9DF, London UK
Phone: +44. 2039361227
Mail: office.uk@hyris.net

Hyris Research Center

Corso Garibaldi, 60
20121 Milano, Italy
Phone: +39. 02.82951302
Mail: office@hyris.net

Hyris Asia Pac

38 Ang Mo Kio Industrial Park 2 #02-07A
569511 Singapore
Phone: +65. 81607207
Mail: office.uk@hyris.net

US West Coast Sales Inquiries

Los Gatos
Mail: office.us@hyris.net

www.hyris.net



Copyright ©2020 Hyris Ltd. All rights reserved.

Reproduction of this document, either in printed or electronics form, is prohibited without the permission of Hyris Ltd.

Information in this document is subject to change without notice. Hyris Ltd assumes no responsibility for any incorrect information that may appear in this document.

Notice:

HYRIS LTD DISCLAIMS ALL WARRANTIES WITH RESPECT TO THIS DOCUMENT, EXPRESSED OR IMPLIED, INCLUDING BUT NOT LIMITED TO THOSE OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. IN NO EVENT SHALL HYRIS LTD BE LIABLE, WHETHER IN CONTRACT, TORT, WARRANTY, OR UNDER ANY STATUTE OR ON ANY OTHER BASIS FOR SPECIAL, INCIDENTAL, INDIRECT, PUNITIVE, MULTIPLE OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH OR ARISING FROM THIS DOCUMENT, INCLUDING BUT NOT LIMITED TO THE USE THEREOF.

Trademarks:

bCUBE and bDATA are registered trademarks of Hyris Ltd.

Microsoft® and Windows® are registered trademarks of Microsoft Corporation. FAM™, HEX™ and ROX™ are a trademark of Thermo Fisher Scientific or its subsidiaries in the U.S. and certain other countries.

Technical Support Contacts:

Hyris Ltd – Italian Branch

Corso Garibaldi 60, 20121 Milano (MI) – ITALY

Phone: +39 02 82951302

Mail: support@hyris.net

Website: <https://www.hyris.net/>

Index

Hyris bCUBE 2.0	4
Box Opening.....	4
Device Overview	5
Status LEDs.....	6
First set up.....	8
PC settings.....	8
bCUBE 2.0 Connection	8
Interacting bCUBE 2.0: bAPP and bPANEL	10
bAPP interface.....	10
Access the bAPP.....	10
Create a New Analysis: general information	11
Create a new analysis: the cartridge	13
Analysis details	15
bPANEL 2.0 interface	18
Installation	18
Main view	18
Settings: Log in/Sync and bCUBE 2.0 connection.....	19
New analysis creation.....	20
Run the analysis.....	21
Old analyses view	21
Flow Chart	23
Appendix-1: Set bCUBE 2.0 Wi-Fi from bCUBE's menu	24
Appendix-2: Custom Recipe	25
Appendix-3: Cartridge Preparation	26
Load samples.....	26
Load/Unload the cartridge to/from bCUBE 2.0	26

Hyris bCUBE 2.0


Box Opening

The following items will be found in the provided box:

- **bCUBE 2.0** (Figure 1a)
- **Ethernet cable** (Figure 1b)
- **Power Supply** including European and North American power cords (Figure 1c).


Characteristics:


- **Input:** 100 ~ 240 V c.a. 1.4A 50/60Hz
- **Output:** 12V DC 5A
- **Power:** 60W
- **USB stick** with Hyris bPANEL software installer and User Manuals (Figure 1d)



Note: Please check the presence of all these items. Remove the items carefully from the shipping box and inspect them for any external damage. If any of the parts is missing or damaged, contact the manufacturer prior to installation.

Device Overview









Description of the features highlighted in the pictures above:


- **Access Port Slider:** opened to insert/remove the cartridge inside bCUBE 2.0. To be kept closed during operations.
- **Access port:** insert the cartridge here. Always push the cartridge up to the end of the port.
- **Thermal Block:** precisely controls the temperature of the cartridge and contained samples. Heating/Cooling is obtained with a Peltier element.
- **Detection block:** features high brightness LEDs with dichroic filters for fluorophores excitation. Detection is obtained thanks to a CMOS sensor with dichroic filters matched to fluorophores emission wavelengths.
- **Top venting holes:** allows air exchange between the inside of the **Instrument** and the ambient
- **Power supply Plug:** insert here the 12 VDC adapter jack
- **Power-on/Power-off button:** press shortly to wake up the instrument from Stand-By. When the machine is running, long press (about 5 seconds) to safely power down the bCUBE
- **Ethernet Plug:** insert here the Ethernet cable for Windows PC / Network communication
- **Status LEDs:** Give information about the status of the instrument


Status LEDs

The status of the bCUBE and its functioning are communicated to the user by the different colour combinations of the status LEDs on the top of the device.


- **bDATA LED:** corresponding to the  symbol, it indicates the connection status to Hyris server.
 Ready colour: green 
 When Ready colour is on, bCUBE 2.0 is connected to bDATA
- **Thermal LED:** corresponding to the  symbol, it indicates if the cartridge is hot or not.
 Ready colour: orange 
 When Ready colour is on, the internal temperature of the bCUBE 2.0 is not hot. If


cartridge is inserted, it can be safely handled

- **Connection LED:** corresponding to the  symbol, it indicates the internet connection state of the bCUBE 2.0




Ready colour: blue 




When ready colour is on, bCUBE 2.0 is connected to a network



For a graphical representation, the  symbol (from left to right: bDATA, Thermal and Connection LEDs) will be displayed in this document when colour code explanation is required.

Some examples are reported in the following table:

LEDs Colours	DESCRIPTION	INTERPRETATION
	<ul style="list-style-type: none"> - bDATA LED steady green - Thermal LED steady orange: - Connection LED steady blue: 	<ul style="list-style-type: none"> - bDATA connection OK - Cartridge not hot - Network connection OK
	<ul style="list-style-type: none"> - bDATA LED steady green - Thermal LED blinking orange and red - Connection LED steady blue 	<ul style="list-style-type: none"> - bDATA connection OK - WARNING: Cartridge hot - Network connection OK
	<ul style="list-style-type: none"> - bDATA LED blinking green and yellow - Thermal LED steady orange 	<ul style="list-style-type: none"> - Missing bDATA connection - Cartridge not hot - Missing network connection

	<ul style="list-style-type: none"> - Connection LED blinking red and yellow 	
	<ul style="list-style-type: none"> - bDATA LED blinking green and yellow - Thermal LED steady orange - Connection LED steady blue 	<ul style="list-style-type: none"> - Missing bDATA connection - Cartridge not hot - Network connection OK. bCUBE 2.0 can be used locally
	<ul style="list-style-type: none"> - bDATA LED multicolour - Thermal LED multicolour - Connection LED multicolour 	<ul style="list-style-type: none"> - Experiment is running on bCUBE 2.0
	<ul style="list-style-type: none"> - bDATA LED off - Thermal LED off - Connection LED off 	<ul style="list-style-type: none"> - bCUBE 2.0 off

Note: not all the Status LEDs colour combinations are shown. For further information, please read the Owner's Manual.

First set up


PC settings





PC has to be connected to a stable internet connection.


bCUBE 2.0 Connection


In order to properly work, the following steps have to be respected:


1. Make sure the bCUBE 2.0 is set on a flat and stable surface
2. Make sure to have correctly provided the condition explained in the [PC Set-up](#) section
3. Connect one end of the Ethernet cable to the bCUBE 2.0 and the other to the PC according the following configuration





4. **Power** on the PC. Wait until PC is ready
5. **Connect** the PC to internet. Make sure internet connection is stable
6. Power on the bCUBE 2.0. Status LEDs colours: 
7. bCUBE will start booting. Status LEDs colours: 
8. Wait until Connection Status LED colour is steady blue:  or 
 If LEDs colours are different from the above mentioned, try to repeat the steps. Otherwise, check the Owner's Manual
9. Access the bAPP (see the [Access the bAPP](#) section) and go to the **bCUBEs** section. Select the desired bCUBE by checking the Serial number of the machine (see the Owner's manual). The **"Connected"** flag should be a green icon, indicating that the bCUBE is connected to a network. An action menu should be show



10. Click in the  button to access the **"Local settings page"**, then go to the **"Network"** section. Select one of the **"Available WiFi Networks"** and enter the related password, then click on **OK**. See the Owner's manual to perform **WPS** connection to the router or to manually add a network. The saved Wi-Fi should immediately appear in the **"Saved WiFi Networks"** column.



11. When all the desired Wi-Fi networks are saved, go back to the bAPP and unplug the Ethernet cable from the bCUBE. After some seconds, the Status LEDs should show the Hyris Ready colours . If not, check the troubleshooting section on the Owner's manual.

Note: Status LEDs may display different colours while bCUBE 2.0 is looking for network connection. If no connection is detected, the following LEDs colours may be displayed after a minute of waiting time: . If the problem persists, see the Owner's Manual.

Interacting bCUBE 2.0: bAPP and bPANEL


bCUBE 2.0 can be controlled and monitored with two Hyris tools: bAPP and bPANEL.

- **bAPP**: it's a web app software and there's no need to install it on the PC. Ready-to-use recipes or custom recipes (see dedicated section) can be used to create and run analyses on the available devices according to the account permissions. Account and devices management are possible by this application. It can be run on PC or Smartphone. **Internet connection is mandatory.**
- **bPANEL**: this software interface is designed for **offline** usage of the bCUBE device, while also providing internet data sharing when connection is available. bPANEL 2.0 can start new analyses by using the already existing recipes (created on bAPP).

bAPP interface


Access the bAPP

1. Make sure to have properly followed the instruction described in the [bCUBE 2.0 Connection](#) section.

If bCUBE 2.0 is ready for operations, open the internet browser (Google Chrome  browser recommended) on the PC and type <https://bapp.hyris.net> in the address bar. Then press enter to access the bAPP.

2. The **Home** window will be shown. The screen is mainly divided in two sections:
 - **Main Menu**: list of available bAPP sections:
 - **Home**: Welcome Window. Log-in credentials required when starting a new session on bAPP
 - **Analyses**: all the available analyses are listed in this section. New analyses can be created
 - **bKITs**: bKITs are reagent sets used to analyse a specific genetic sequence. The bKITs developed and sold by Hyris Ltd are listed in this section
 - **Recipes**: a "Recipe" is a detailed list of steps and settings that the bCUBE has to execute while performing an analysis. New recipes can be customized from this menu (see [Appendix-2: Custom Recipe](#)). All the available recipes are listed here
 - **Swarms**: "Swarm" is used to define a collector of analyses and recipes. User can have access to multiple swarms according to his permissions. Visible Swarms are listed in this section
 - **Users**: list of available users and relative permissions (see Owner's Guide for further information)
 - **bCUBEs**: list of available bCUBEs with relative extra information (see Owner's Guide for further information)
 - **Certificates**: list of requested certificates.

- **Tab View Area:** content is related to the selected Main Menu section



3. After the Main Window is displayed, **Login** is required.
 Insert your email and password, which are corresponding to the **credentials** used to register the user to the Hyris Ltd web site or the bAPP. Enable the “*Remember me*” option to keep the user logged in.
 In order to sign up, push the **SIGN UP** button and follow the steps.

In case of forgotten password, press on **FORGOT PASSWORD?** and insert the user mail. Then check your mail box for further instructions.


Create a New Analysis: general information

Go to the “**Analyses**” section from the **Main Menu**. The already existing analyses will be listed in a table.

In order to create a new analysis, click on the **+** button.
 The user will be guided through few steps to correctly fill all the requested information.
 The first step refers to the general information of the analysis:


- **Name:** name of the analysis
- **Description** (Optional): brief description of the experiment
- **Swarm:** select the Swarm to which the analysis will be linked to
- **Recipe:** choose a recipe among the ones related to the selected Swarm.
- **bKITS:** show available bKITS (associated to the relative global recipe) instead of Swarm recipes

Click on **CONTINUE** to proceed to the following step




Custom recipe choice

In case of disabled “bKITS” option:

1. Select the desired **Swarm** among the list. In case only one available swarm for the user, it will be automatically selected. After the selection, the relative swarm recipes are loaded.
2. Select the desired **Recipe** among the list. Click on the  button to see the recap of the thermal protocol relative to the selected recipe.

bKIT choice

In the new/edit analysis page, after enabling bKITS **[A]**, the following parts have to be filled:



- A. bKITS Flag:** if bKITS are enabled, fields **[B]**, **[C]**, **[E]** and **[F]** are shown
- B. bKITS categories (mandatory):** the available bKITS depend on the selected category (examples: **Probiotics, Plant Health, Botanicals ...**)
- C. bKIT part number (mandatory):** list of available bKITS according to the selected category. The bKIT part number is reported on the bKIT package as shown below.




- D. **QR Code Scanner:** if the desired bKIT is provided with dedicated QR code, click on this button and scan it. The QR code could be placed on the bKIT box and / or in the provided datasheet. In this case, the fields **[B]** and **[C]** are automatically set.
- E. **Global Recipe:** this is a read only field, which shows the name of the Global Recipe associated to the selected bKIT. It's automatically filled once **[C]** is selected
- F. **Datasheet/DNA extraction method download buttons:** click on the desired button to download a .pdf file of the desired document


Note: Each **VERSION** of the bKIT is associated with a specific **Global Recipe**. Verify that the selected **bKIT** and its **Version** match the part number and version reported on the package.


Create a new analysis: the cartridge


Select the desired cartridge among the available ones (for example: "4x4 samples").

1. The wells information is filled while compiling the selected recipe.



2. Wells can be deleted (press the  linked to the desired well), modified or added. In this last case:



- a) Click on the desired well on the cartridge model on the right, according to the loaded sample
- b) A dedicated window will appear
- c) Select the **Sample Type** from the drop-down menu (Mandatory)
- d) Enter the **Sample Name** (Mandatory)
- e) Choose the **Well Colour**
- f) Click on  button






The dialog box titled "Define well nr. 2" contains the following elements:


- A dropdown menu labeled "Select well type" with a downward arrow.
- A text input field labeled "Name".
- A color selection grid with 48 small colored squares arranged in 6 rows and 8 columns.
- A large green square to the right of the color grid.
- At the bottom right, two buttons: "CANCEL" and "SAVE".


Other selection modalities are available for desktop mode:

- **Multiple selection of contiguous wells:** hold down the left button of the mouse. A cross cursor will appear. Holding down the button, drag the mouse on the cartridge to select many contiguous wells. Click on the  button to open the "edit well" window.
- **Multiple selection of non-contiguous wells:** hold down the **ctrl** button (a pointer cursor will appear) and click on the desired wells. Click on the  button to open the "edit well" window.

You can then add a new well by repeating this procedure or click on .

If the "**Launch now**" option is selected, a new window will open. Select the desired bCUBE among the available devices. The selected bCUBE status LEDs will blink. You can make the Status LEDs blinking again by clicking on the  button. A recap of the thermal protocol that will be executed is visible by clicking the  button. *Cartridge Loading Wizard* can also be shown, and cartridge .PDF file can be downloaded. Click on  to immediately launch the analysis. The live analysis view page will be shown.



The status LEDs of the device will start to show multiple and changing colours . Otherwise, check the Owner's manual for further information.

If the **"Save for later"** option is selected, the user will be brought back to the Analyses list view.




Analysis details

GENERAL tab

This section shows the cartridge temperature measured by the instrument during the experiment and some information about the operating bCUBE 2.0 (for example, its name and software version) or the analysis.

Each of these parameters are updated in real time while analysis is running and bCUBE 2.0 is connected to the internet.

If the analysis state is **"Running"**, the user can send commands to the operating bCUBE 2.0:

-  **ABORT** : the running analysis will be immediately stopped
-  **EXTEND (5 ...)** : the extend button will add 5 cycles/minutes to the current step. This operation is not allowed when the "Melting Curve - VAL" step is running.
-  **SKIP** : this command instantly ends the current step in order to start with the following one.

Each of the previous actions will be tracked both in the analysis report and in the step preview on the **Tab View Area**.



WARNING


Make sure the previews command has been executed by bCUBE 2.0 before sending a new command.

STEPS tab

All analysis steps are listed in this tab.

Each step data can be consulted individually by clicking on the correspondent name on the list.

The step view will be expanded, showing extra data, such as the fluorescence/time graphic, which is updated in real time while the analysis is "Running" and bCUBE 2.0 is connected to the internet.



If the analysis is completed, a computed parameter table will be added next to the graphic (if specified in the recipe).


Select parameter: TTS

Search Table


Sample	Type	Channel	Value
Positive control	PosCtrl	FAM	0:00:56
Unknown 1	Sample	FAM	0:00:55
Unknown 2	Sample	FAM	0:00:55
Negative control	NegCtrl	FAM	
Unknown 3	Sample	FAM	0:01:08
Unknown 4	Sample	FAM	0:01:07
Unknown 5	Sample	FAM	0:01:08
Unknown 6	Sample	FAM	

RESULTS Tab








In this tab a graphic representation of the selected cartridge is shown. After the analysis is completed, each sample is marked with a label indicating the result of the experiment (if specified in the recipe).



Action button

Click on the  button in the tab section to open the action menu.

Different actions are available depending on the analysis status:

-  **Generate PDF** button will immediately start the download of the report (.pdf file) of the current analysis
-  **Generate .XLSX** button will download an Excel readable file with all the analysis information and results.
-  **Edit Samples** button allows the user to change some wells information, such as the wells names or types.
-  **Recalculate results** allows the user to change some settings of each step and to recompute the analysis results accordingly.
-  **EXTEND** the extend button will add 5 cycles/minutes to the current step. This operation is not allowed when the "Melting Curve - VAL" step is running.
-  **ABORT** button: the running analysis will be immediately stopped
-  **SKIP** button: this command instantly ends the current step in order to start with the following one.

bPANEL 2.0 interface

Installation

Open the bPANEL 2.0 installer located in the provided USB.

- 1) Insert into an available USB port the Hyris USB Stick containing bPANEL installer.
- 2) Launch the file **Setup.msi**, located inside **bPANEL** folder.
- 3) On the welcome screen, click on **Next**.





Figure 1: bPANEL installation - Welcome screen

- 4) On the next page, carefully read the licence agreement. Tick the check-box and click **Install** to continue.
- 5)



- 6) Wait for the installation process to end. Click on **Finish** on the last page to complete the installation.

No driver installation is required for using bcUBE 2.0.

Main view

Make sure all the steps described in [bcUBE 2.0 Connection](#) have been correctly performed.

Open the bPANEL 2.0 program. The main view will be shown:

- **Command Buttons:** list of command to send to the connected bcUBE
- **Tab selection:** each tab provides different information about the loaded analysis or the user settings. The analysis related tabs are updated in real time while the selected analysis is running and the correspondent bcUBE is connected


- **Tab View Area:** shows the content of the selected tab


Figure 2: bPANEL Overview


Settings: Log in/Sync and bCUBE 2.0 connection

Go to “Settings” tab:

1. Click on



2. Insert your **email** and **password**, which are corresponding to the **credentials** used to register the user to the Hyris Ltd web site (same as used credential to access the bAPP. See [Access the bAPP](#) section for further information). Then click “OK”




Login

bData Username: TestUser@demo.net

bData Password: *****

Ok Cancel

- Click on **Search**. After the operation completion, the available bCUBE devices will be listed in the dedicated drop-down menu. Select the desired one



Protocol Settings



bCube IP/Name: XXX.XXXX.XX.XX.XX Search

bCube Host Name: XX.XXXXXXXXXXX


bData Username: TestUser@demo.net


Last Sync: 26/09/2019 11:44:29

Log Out Log In Sync

- Click on the  button to connect to the selected device. After connection, the button will turn green 

New analysis creation

- Click on the  button
- Insert the **Analysis Name** (Mandatory)
- Select the desired **Swarm** among the available ones (Mandatory)
- Click on **Global recipes**
- Select the desired **Recipe**
- Click on **Ok** button.



Create new Analysis

Analysis name: DEMO_ANALYSIS

Select Cartridge type: 16wells


Select Swarm: Test

☐ Swarm recipes ☒ Global recipes


Select Recipe: Panax spp. 1.1

Ok Cancel



7. The new analysis will be created and tabs updated accordingly.
The wells information are filled while compiling the selected recipe downloaded from bDATA.



8. Wells can be deleted, modified or added. In this last case:
 - a. Click on the desired well on the cartridge model on the right, according to the loaded sample
 - b. Select the **Sample Type** from the drop-down menu (Mandatory)
 - c. Enter the **Sample Name** (Mandatory)
 - d. Choose the **Sample Colour**




Run the analysis


1. Load the cartridge to the selected bCUBE 2.0 (see section: [Cartridge Preparation](#)).
2. Run the analysis by clicking on the  command button. The analysis will be uploaded to the selected bCUBE 2.0. If no errors occur, the experiment will immediately start, while the status LEDs of the device will start to show multiple and changing colours.
3. The analysis parameters and results will be automatically shown in specific tabs in real time.
4. Wait for "Ready Colours"
5. Generate a PDF report file by clicking on the  command button, and then selecting **"Build/Rebuild PDF Report"**.
6. Unload the cartridge

Old analyses view

Old analyses can be consulted by opening them from PC, bCUBE or online bDATA folders.



Click on the  command button and select the desired folder location among “**LOCAL**”, “**bCUBE**” and “**bDATA**”. Click on “**OK**” after the analysis selection. The analysis data will be loaded to bPANEL and results can be consulted. The opened analysis can’t be modified.


Flow Chart



Appendix-1: Set bCUBE 2.0 Wi-Fi from bCUBEs menu

If the bDATA Status LED is showing steady green colour, bCUBE 2.0 Wi-Fi can also be configured from “**bCUBEs**” section of the bAPP.

1. Access the bAPP and Log In as shown in the [PC: bAPP Interface](#) section
2. Click on “**bCUBEs**” from the main menu: a list of the available devices will appear. The available devices could be different among users of the same company depending on their permissions.
3. Select the desired device and press the  button.
4. Select the desired LAN and set the Wi-Fi password, then click on .



Appendix-2: Custom Recipe

New recipes can be customized on the bAPP according to user's permissions. If recipe customization is not allowed by user's permissions, global recipes provided by Hyris Ltd or already existing company recipes can be used for new analyses.

1. Click on "**Recipes**" in the left side **Main Menu**: the list of the available recipes will be shown.

Recipes				+	↺	Columns ▼
<input type="text" value="Search"/>						All Columns ▼
↕ Date	🔍 Name	🔍 Swarm	🔍 Description			
○ 2019-09-18 15:26	ExampleRecipe	Test				

2. Click on the **+** button to access to the new recipe customization form.
Remember that **Recipe name** and selected **Swarm** are mandatory.
3. Follow the steps and click on **FINISH** button.
4. The new recipe will be added to the recipe list

Appendix-3: Cartridge Preparation

Load samples

In order to load the samples for the experiment:

1. Place the cartridge face up on a flat and stable surface
2. Put the sample in one or more of the cartridge wells as indicated in the reference Recipe.
3. Seal the cartridge with the provided aluminium foils



Load/Unload the cartridge to/from bCUBE 2.0

To load the cartridge to the bCUBE 2.0:

1. Open the bCUBE 2.0 Access Port slider
2. Insert the cartridge by pushing it through the Access Port by keeping it face up holding it from the handle till it stops
3. Close the access port slider

To unload the cartridge:

1. Make sure the cartridge is not hot as indicated by the Steady orange colour of the Thermal Status LED.
2. Open the access port slider
3. Gently pull the cartridge from the handle
4. Close the access port slider