



bKIT *Lactobacillus acidophilus HA-122*

Real-Time PCR assay

Hyris Ltd

HYRIS Headquarters Lower Ground Floor, One George Yard, EC3V 9DF, London, UK Phone: +44.2036082968 Mail: office@hyris.net

HYRIS Research Center Corso Garibaldi 60. Milano 20121, Italy Phone: +39.02.82951302 administrator@hyris.net Mail:

HYRIS Asia Pac Block 998 Toa Payoh North #06-06 Toa Payoh North Industrial Estate Singapore 318993, Singapore Phone: +65.8160.7207 Mail: office@hyris.net

www.hyris.net



HYRIS Headquarters

Lower Ground Floor, One George Yard, EC3V 9DF, London, UK Phone +44 2036082968 Mail: office@hyris.net

HYRIS Research Center Corso Garibaldi 60

Milano 20121, Italy +390282951302 Phone: Mail: administrator@hyris.net

HYRIS Asia Pac

Block 998 Toa Payoh North #06-06 Toa Pavoh North Industrial Estate Singapore 318993, Singapore Phone: +65.8160.7207 Mail: office@hyris.net

www.hyris.net

WARNING AND PRECAUTIONS

Do not mix the components of kits with different batch number. This kit is designed to be used by personnel trained to follow correct molecular biology precautions

WARRANTY AND RESPONSIBILITIES

Hyris Ltd guarantees the buyer exclusively concerning the quality of reagents and of the components used to produce the Kits. Hyris Ltd is not responsible and cannot anyway be considered responsible or jointly responsible for possible direct and indirect damages resulting from the utilization of the Kits by the user. The user consciously and under his/her own responsibilities decides for the utilization purposes of the Kits and uses it the way he/she considers most suitable in order to reach his/her goals and/or objectives. Hyris is not responsible for the data resulting

from the use of the Kits, for the utilization that the user independently decides to make of them or for the direct or indirect damages possibly resulting from the disclosure or transmission of the data themselves to third parties under any form or circumstance. This clause automatically accepted by the user when

purchasing the Kits. The patent for performing PCR and Real-Time PCR is held by Hoffmann/La Roche. Authorization to use PCR and Real-Time PCR can be obtained on licence from Hoffmann-La Roche. The product equipment and information included in the Kits equipment and information included in the KIS consist of assembled reagents. The licence and authorisation for PCR and Real-Time PCR use are not included in the KIS. The user is responsible for setting prefixed goals, choosing whether or not to perform the PCR or Real-Time PCR reaction and to apply for register his/her own licence. The patent for performing LAMP is held by Eiken, from which authorization to use LAMP can be obtained. The license for LAMP is not included in the Kits.

The Kits are designed for the services supply Diagnostic applications that are not research use only require a specific licence for PCR and Real-Time PCR or for LAMP use. The Kits may require the use of Taq Polymerase enzyme and/or fluorochromes, often registered as trademark by companies. TagMan® is a trademark of Roche Molecular Systems, Inc. FAM, VIC and ROX are a trademark of Applera Corporation or its affiliate. Cy5 is a trademark of Amersham Biosciences Ltd. The Kits have been internally tested by our quality control. Any responsibility is waived if the warranty of quality control does not refer to the specific Kits. The user is personally responsible for data that he/she will obtain and/or he/she will supply to third parties using these kits. Once the sealed package is opened the user accepts all the conditions without fail; if the package is still sealed the kit can be returned and the user can be refunded. The Kits are intended for Research Use Only

Introduction

Lactobacillus classification trace back to 1901, when, based on biochemical and morphological characteristics, Beijerinck¹ M.W proposed the genus. Using the same approach Moro², Hansen and Mocquot³ proposed the species Lactobacillus acidophilus. Nowadays, many efforts focus on the correlation between taxonomic classification with traditional procedures and DNA molecular methods. Consistently with this trend, traditional culture approaches are increasingly assisted by DNA molecular methods⁴. Among these, Real-Time PCR emerged for its sensitivity, rapidity, reliability, specificity and repeatability making it a well-established method for the detection, quantification, and typing of different microbial agents in the areas of clinical and veterinary diagnostics and food safety⁵.

(1) BEIJERINCK (M.W.): Sur les ferments lactiques de l'industrie. Archives Néerlandaises des Sciences Exactes et Naturelles (Section 2), 1901, 6, 212-243.

(2) MORO (E.): Über den Bacillus acidophilus n. sp. Jahrbuch fur Kinderheilkunde und physiche Erziehung, 1900, 52, 38-55.

(3) HANSEN (P.A.) and MOCQUOT (G.): Lactobacillus acidophilus (Moro) comb. nov. International Journal of Systematic Bacteriology, 1970, 20, 325-327.

(4) Mianzhi Y, Shah NP. Contemporary nucleic acid-based molecular techniques for detection, identification, and characterization of Bifidobacterium. Crit Rev Food Sci Nutr. 2017 Mar 24;57(5):987-1016. doi: 10.1080/10408398.2015.1023761. Review.PubMed PMID: 26565761.

(5) Kralik P, Ricchi M. A Basic Guide to Real Time PCR in Microbial Diagnostics: Definitions, Parameters, and Everything. Front Microbiol. 2017 Feb 2;8:108. doi: 10.3389/fmicb.2017.00108. eCollection 2017. Review. PubMed PMID: 28210243; PubMed Central PMCID: PMC5288344.

Principle

SYBR® Green Real-Time PCR (qPCR) assay for the detection of Lactobacillus acidophilus HA-122. The product is intended for research purpose only.

NHPRA validation

In the validation trials performed by NHPRA (Natural Health Product Research Alliance) the following strains were tested: Lactobacillus acidophilus HA-122, Lactobacillus casei HA-108, Lactobacillus paracasei HA-196, Lactobacillus paracasei HA-274, Lactobacillus acidophilus R0418. Moreover, assay performances were assessed in mixtures containing the DNA of the strains listed above. All DNA solutions tested were normalized to the concentration of $1 \text{ ng/}\mu\text{L}$ before use. All target and non-target DNA sample solutions were successfully classified. For more details, contact us at support@hyris.net.

bKIT Lactobacillus acidophilus HA-122 packaging

Part Number: bKTPR-LAHA122.01-50

qPCR Master Mix (1 tube)	50 tests
Positive Control (1 tube)	10 tests
Negative Control (1 tube)	10 tests

Part Number: bKTPR-LAHA122.01-100

qPCR Master Mix (2 tubes)	2 x 50 tests
Positive Control (1 tube)	20 tests
Negative Control (1 tube)	20 tests

Storage

-20°C. Avoid prolonged exposure to light and repeated freeze and thaw cycles.

Shelf life

If the bKIT is correctly stored, at constant-temperature freezer, its performance is guaranteed until the shelf life indicated on the tubes.



HYRIS Headquarters Lower Ground Floor, One George Yard,

EC3V 9DF. London, UK +44.2036082968 Phone: Mail: office@hyris.net

HYRIS Research Center Corso Garibaldi 60,

Milano 20121, Italy Phone: +39.02.82951302 Mail: administrator@hyris.net

HYRIS Asia Pac

Block 998 Toa Pavoh North #06-06 Toa Payoh North Industrial Estate Singapore 318993, Singapore Phone: Mail: +65.8160.7207 office@hyris.net

www.hyris.net

WARNING AND PRECAUTIONS

Do not mix the components of kits with different batch number. This kit is designed to be used by personnel trained to follow correct molecular biology precautions

WARRANTY AND RESPONSIBILITIES

Hyris Ltd guarantees the buyer exclusively concerning the quality of reagents and of the components used to produce the Kits. Hyris Ltd is not responsible and cannot anyway be considered responsible or jointly responsible for possible direct and indirect damages resulting from the utilization of the Kits by the user. The user consciously and under his/her own responsibilities decides for the utilization responsibilities decides for the utilization purposes of the Kits and uses it the way he/she considers most suitable in order to reach his/her goals and/or objectives. Hyris is not responsible for the data resulting from the use of the Kits, for the utilization that

the user independently decides to make of them or for the direct or indirect damages possibly resulting from the disclosure or transmission of the data themselves to third parties under any form or circumstance. This clause automatically accepted by the user when

purchasing the Kits. The patent for performing PCR and Real-Time PCR is held by Hoffmann/La Roche. Authorization to use PCR and Real-Time PCR can be obtained on licence from Hoffmann-La Roche. The product equipment and information included in the Kits equipment and information included in the KIS consist of assembled reagents. The licence and authorisation for PCR and Real-Time PCR use are not included in the KIts. The user is responsible for setting prefixed goals, choosing whether or not to perform the PCR or Real-Time PCR reaction and to apply for register his/her own licence. The patent for performing LAMP is held by Eiken, from which authorization to use LAMP can be obtained. The license for LAMP is not included in the Kits.

The Kits are designed for the services supply Diagnostic applications that are not research use only require a specific licence for PCR and Real-Time PCR or for LAMP use. The Kits may require the use of Taq Polymerase enzyme and/or fluorochromes, often registered as trademark by companies. TagMan® is a trademark of Roche Molecular Systems, Inc. FAM, VIC and ROX are a trademark of Applera Corporation or its affiliate. Cy5 is a trademark of Amersham Biosciences Ltd. The Kits have been internally tested by our quality control. Any responsibility is waived if the warranty of quality control does not refer to the specific Kits. The user is personally responsible for data that he/she will obtain and/or he/she will supply to third parties using these kits. Once the sealed package is opened the user accepts all the conditions without fail; if the package is still sealed the kit can be returned and the user can be refunded. The Kits are intended for Research Use Only

Additional material/reagents required

- DNA extraction tools and reagents
- Nuclease-free water .
- Gloves .
- . Pipettes
- bCUBE[®] instrument or other Real-Time PCR instrument (*) with filters calibrated for SYBR[®] • Green.
- bCUBE[®] sample loading cartridge or, if using other Real-Time PCR instrument, samples loading support according to the instrument specifications.
- (*) This assay was especially developed to be used in association with the bCUBE[®] instrument, available from Hyris Ltd, but can be used also with any other compatible thermal cycler.

DNA extraction

In the validation trials performed by NHPRA the DNA was extracted using NucleoSpin® Food (MACHEREY-NAGEL) and normalized to the concentration of 1 ng/µL.. For more details, contact us at support@hyris.net.

Reaction Set-Up

- a. Thaw all the bKIT components by placing the tubes on ice.
- b. Gently mix the tubes content by swirling the tubes.
- Spin the tubes to let the content down. c.
- d. In new tubes, one for each sample, including the Negative Control and the Positive Control of the bKIT, prepare the Reaction Mix as shown in the table below:

Components	Volume
DNA sample (normalized to the concentration of 1 $ng/\mu L$) or Positive Control/Negative Control	2 μΙ
qPCR Mastermix	18 µl
Total Volume	20 µl

Cartridge set-up

The procedure described is for the bCUBE® cartridge, but, if using a different Real-Time PCR instrument, the same procedure can be adopted for other loading sample supports with minor modifications.

1. Samples set-up

Samples of the following types must be prepared to be loaded on the cartridge: Positive Control for Lactobacillus acidophilus HA-122.

Negative Control for Lactobacillus acidophilus HA-122. Sample(s) to be tested.

2. Cartridge Loading

- Load the sample prepared as described in the previous section. a.
- Carefully seal the cartridge with adhesive film in order to avoid any contamination. b.
- Load the cartridge onto the bCUBE[®], then start the run. c.

Method set-up

Set up the run method using the following conditions, depending on the instrument you use.



HYRIS Headquarters Lower Ground Floor, One George Yard,

EC3V 9DF, London, UK Phone +44.2036082968 Mail: office@hyris.net

HYRIS Research Center Corso Garibaldi 60, Milano 20121, Italy

+39.02.82951302 Phone: Mail: administrator@hyris.net

HYRIS Asia Pac Block 998 Toa Payoh North #06-06 Toa Payoh North Industrial Estate Singapore 318993, Singapore Phone: +65.8160.7207 Mail: office@hyris.net

www.hyris.net

WARNING AND PRECAUTIONS

Do not mix the components of kits with different batch number. This kit is designed to be used by personnel trained to follow correct molecular biology precautions

WARRANTY AND RESPONSIBILITIES

Hyris Ltd guarantees the buyer exclusively concerning the quality of reagents and of the components used to produce the Kits. Hyris Ltd is not responsible and cannot anyway be considered responsible or jointly responsible for possible direct and indirect damages resulting from the utilization of the Kits by the user. The user consciously and under his/her own responsibilities decides for the utilization responsibilities decides for the utilization purposes of the Kits and uses it the way he/she considers most suitable in order to reach his/her goals and/or objectives. Hyris is not responsible for the data resulting from the use of the Kits, for the utilization that

the user independently decides to make of them or for the direct or indirect damages possibly resulting from the disclosure or transmission of the data themselves to third parties under any form or circumstance. This clause automatically accepted by the user when

purchasing the Kits. The patent for performing PCR and Real-Time PCR is held by Hoffmann/La Roche. Authorization to use PCR and Real-Time PCR can be obtained on licence from Hoffmann-La Roche. The product equipment and information included in the Kits equipment and information included in the KIS consist of assembled reagents. The licence and authorisation for PCR and Real-Time PCR use are not included in the KIS. The user is responsible for setting prefixed goals, choosing whether or not to perform the PCR or Real-Time PCR reaction and to apply for register his/her own licence. The patent for performing LAMP is held by Eiken, from which authorization to use LAMP can be obtained. The license for LAMP is not included in the Kits.

The Kits are designed for the services supply Diagnostic applications that are not research use only require a specific licence for PCR and Real-Time PCR or for LAMP use. The Kits may require the use of Taq Polymerase enzyme and/or fluorochromes, often registered as trademark by companies. TagMan® is a trademark of Roche Molecular Systems, Inc. FAM, VIC and ROX are a trademark of Applera Corporation or its affiliate. CyS is a trademark of Amersham Biosciences Ltd. The Kits have been internally tested by our quality control. Any responsibility is waived if the warranty of quality control does not refer to the specific kits. The user is personally responsible for data that he/she will obtain and/or he/she will supply to third parties using these kits. Once the sealed package is opened the user accepts all the conditions without fail; if the package is still sealed the kit can be returned and the user can be refunded. The Kits are intended for Research Use Only

On the bCUBE® 1.

- Login on the bAPP. a.
- Set-up "New Analysis" and Select the "Lactobacillus acidophilus HA-122 1.x" from the b. "Global recipes" list.
- Specify the "Well types" for each of the loaded sample as follows (Fig. 1): c. "PosCtrl" for the well loaded with Lactobacillus acidophilus HA-122 Positive Control. "NegCtrl" for the well loaded with Lactobacillus acidophilus HA-122 Negative Control. "Sample" for the wells loaded with samples under analysis.





An example of cartridge set-up on the bAPP for one replicate of a sample to be analyzed is shown.

On a compatible Real-Time PCR instrument 2.

Please, contact us for the protocol set-up on the instrument.

Reading the results

1. On the bCUBE®

The presence of the target Lactobacillus acidophilus HA-122 in the Positive Control or in the a. sample under analysis will generate an amplification curve (Fig. 2a) and a melting curve with a specific melting peak (Fig. 2b).



In the plots, the amplification curve (Fig. 2a) and the specific melting peak (Fig. 2b) of a Lactobacillus acidophilus HA-122 containing sample (A), a Positive Control (B), and the Negative Control (C) are shown.

b. At the end of analysis each well will be labelled depending on the "Well type" as described in the table below and samples classification will be shown on the pdf report of the analysis (Fig. 3).

Well type	Possible labels	
Positive Control (PosCtrl)	ОК	ко
Label meaning	Amplification curve and specific melting peak present	Amplification curve and or specific melting peak absent



HYRIS Headquarters Lower Ground Floor, One George Yard, EC3V 9DF, London, UK

+44.2036082968 Phone: Mail: office@hyris.net

HYRIS Research Center Corso Garibaldi 60,

Milano 20121, Italy +39.02.82951302 Phone: Mail: administrator@hyris.net

HYRIS Asia Pac Block 998 Toa Payoh North #06-06 Toa Payoh North Industrial Estate Singapore 318993, Singapore Phone: +65.8160.7207 office@hyris.net Mail:

www.hyris.net

WARNING AND PRECAUTIONS

WARRANTY AND RESPONSIBILITIES

biology precautions

Do not mix the components of kits with different

batch number. This kit is designed to be used by personnel trained to follow correct molecular

Hyris Ltd guarantees the buyer exclusively concerning the quality of reagents and of the components used to produce the Kits. Hyris Ltd is not responsible and cannot anyway be considered responsible or jointly responsible for possible direct and indirect damages resulting from the utilization of the Kits by the user. The user consciously and under his/her own responsibilities decides for the utilization

responsibilities decides for the utilization purposes of the Kits and uses it the way he/she considers most suitable in order to reach his/her goals and/or objectives. Hyris is not responsible for the data resulting from the use of the Kits, for the utilization that

the user independently do die studie autor that the user independently docides to make of them or for the direct or indirect damages possibly resulting from the disclosure or transmission of the data themselves to third parties under any form or circumstance. This clause is undependently economic to the the undependent

automatically accepted by the user when purchasing the Kits. The patent for performing PCR and Real-Time PCR is held by Hoffmann/La Roche. Authorization to use PCR and Real-Time PCR can be obtained on licence from Hoffmann-La Roche. The product equipment and information included in the Kits equipment and information included in the KIS consist of assembled reagents. The licence and authorisation for PCR and Real-Time PCR use are not included in the KIS. The user is responsible for setting prefixed goals, choosing whether or not to perform the PCR or Real-Time PCR

reaction and to apply for register his/her own licence. The patent for performing LAMP is held by Eiken, from which authorization to use LAMP can be obtained. The license for LAMP is not

included in the Kits.

bKIT Lactobacillus acidophilus HA-122 – Real-Time PCR assay

Well type	Possible labels	
Negative Control (NegCtrl)	ОК	ко
Label meaning	Amplification curve and specific melting peak absent	Amplification curve and or specific melting peak present

Well type	Possible labels		
Sample	Present	Absent	Indeterminate
Label meaning	Lactobacillus acidophilus HA-122 is present in the sample	Lactobacillus acidophilus HA-122 is absent from the sample	The test is not conclusive and should be repeated (**)

(**) If the "Indeterminate" classification persists, contact us at support@hyris.net.

Results for target Lactobacillus acidophilus HA-122			
Positive control	(PosCtrl)	ОК	
Unknown sample	(Sample)	Present	
Negative control	(NegCtrl)	ОК	

Fig.3. Analysis results table

The results table, as reported in the pdf report of the analysis, is shown.

2. On a compatible Real-Time PCR instrument

Please, contact us for results interpretation.

Troubleshooting

1. Results show no amplification, or anomalous amplification curves

Possible causes	Corrective actions
Evaporation of the sample due to inadequate sealing of the plate/strips	Repeat the test using the appropriate materials and tools to seal correctly the plate/strips
Consumables are not appropriate for the method	Repeat the test using consumables recommended by the supplier of the Real-Time PCR instrument
The quality of nucleic acid extracted is low	Repeat the extraction step. Ensure that the method of extraction has been performed correctly. In any doubt, contact us.

No amplification curve is observed for the Positive Control 2.

Possible causes	Corrective actions
The Positive Control provided with the assay was	Repeat the test adding the Positive Control.
not added into the reaction well	If the problem persists, contact us.

An amplification curve with a specific melting peak is observed for the Negative Control 3.

Possible causes	Corrective actions
Contamination of the Negative Control or the qPCR	Repeat the test by applying appropriate quality procedures to prevent contamination.
Master Mix with target-positive DNA	Correctly seal the cartridge or plate/strips.
	If the problem persists, contact us.

Document revision April 08th 2019

