

# Certificate



Certificate no.

CU 72181696 01

**License Holder:**

Hyris Ltd.  
Lower Ground Floor  
One George Yard  
London  
EC3V 9DF  
United Kingdom

**Manufacturing Plant:**

[REDACTED]  
[REDACTED]  
[REDACTED]  
20157 Milano, MI  
Italy

Test report no.: USA-AK 50163894 001

Client Reference:

Tested to: UL 61010-1:2012  
CAN/CSA-C22.2 NO. 61010-1-12

Certified Product: Bio-Molecular Analyzer

License Fee - Units

Model Designation: bCUBE 2.0

7

Rated Voltage: 12Vdc (for Unit)

Rated Current: 5A (for Unit)

Special Remarks: This product has also been evaluated and found in compliance with the requirement of UL 61010-2-010:2015, CAN/CSA-C22.2 NO. 61010-2-010:15, UL 61010-2-081:2015, CAN/CSA-C22.2 NO. 61010-2-081:15.

Appendix: 1, 1-4

7

Licensed Test mark:



Date of Issue

(day/mo/yr)

20/07/2018